Certificate of Analysis

InTray™ DM-FungID/Dermatophyte

Product Numbers: 11-593-001, 11-593-002
12-063-001, 12-063-002

Lot Number: 260902B

Expiration Date: 23DEC2028

Test		Specification	Results
Appearance		Clear, Yellow	Clear, Yellow
Microbial load		No Growth	No Growth
pH		5.6 ± 0.1	5.65
Microbiology (ATCC®):		**Colony Appearance**	
T. mentagrophytes	9533	Growth w/Color Change ≤ 7 days	Growth w/ Color Change @ 7 days
T. rubrum	28188	Growth w/Color Change ≤ 7 days	Growth w/ Color Change @ 7 days
M. gypseum	14683	Growth w/Color Change ≤ 7 days	Growth w/ Color Change @ 7 days
C. albicans	60193	Partially Inhibited, Weak Color Change ≥7 days	Partially Inhibited, Weak Color Change @ 7 days
A. brasiliensis	16404	No Growth/Significant Inhibition	Significant Inhibition
S. aureus	25923	Inhibited	Inhibited
E. coli	25922	Inhibited	Inhibited

QC Acceptance Date:05OCT2026

Signature: Megan Slama

Biomed Diagnostics, a DCN Dx brand
3193 Lionshead Ave. Ste. 200, Carlsbad, CA 92010, USA
1 (541) 830-3000, www.biomeddiagnostics.com; info@biomeddiagnostics.com
Doc No: 100-542, Rev. F, Effective Date: (09/2025)